

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Rogowski current sensor
Name and address of the applicant	ALGODUE ELETTRONICA SRL VIA P. GOBETTI 16/F 28014 MAGGIORA NO Italy
Name and address of the manufacturer	ALGODUE ELETTRONICA SRL VIA P. GOBETTI 16/F 28014 MAGGIORA NO Italy
Name and address of the factory	<input type="checkbox"/> Additional information on page 2 ALGODUE ELETTRONICA SRL VIA P. GOBETTI 16/F 28014 MAGGIORA NO Italy
Note: When more than one factory, please report on page 2	
Ratings and principal characteristics	3.3 Vd.c. - 10 mA
Trademark (if any)	SOCOMEK
Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	TF-XXX Series Models TF-40 - TF-80 - TF-120 - TF-200 - TF-300 - TF-600 (see Test Report)
Additional information (if necessary may also be reported on page 2)	<input type="checkbox"/> Additional information on page 2 This CBTC is a modification to CBTC no. IT-18930 dated 2018-04-23 and they must be read in conjunction. Modification introduced: adding of new model TF-40 and modification of applicant and manufacturer address.
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010, IEC 61010-1:2010/AMD1:2016, IEC 61010-2-032:2012 National differences: CA, US
As shown in the Test Report Ref. No. which forms part of this Certificate	AE20-0059038-02

This CB Test Certificate is issued by the National Certification Body

IMQ S.p.A.
Via Quintiliano 43, IT-20138 Milano, Italy